



The accuracy of a 3D printing surgical guide determined by CBCT and model analysis

Boyoung Ma¹, Taeseok Park², Inkon Chun², Kwidug Yun^{1*}

¹Department of Prosthodontics, School of Dentistry, Chonnam National University, Gwangju, Republic of Korea

²DMAX Co. Ltd., Gwangju, Republic of Korea

PURPOSE. The aim of this clinical study was to assess the accuracy of the implants placed using a universal digital surgical guide. **MATERIALS AND METHODS.** Among 17 patients, 28 posterior implants were included in this study. The digital image of the soft tissue acquired from cast scan and hard tissue from CBCT have been superimposed and planned the location, length, diameter of the implant fixture. Then digital surgical guides were created using 3D printer. Each of angle deviations, coronal, apical, depth deviations of planned and actually placed implants were calculated using CBCT scans and casts. To compare implant positioning errors by CBCT scans and plaster casts, data were analyzed with independent samples t-test. **RESULTS.** The results of the implant positioning errors calculated by CBCT and casts were as follows. The means for CBCT analyses were: angle deviation: $4.74 \pm 2.06^\circ$, coronal deviation: 1.37 ± 0.80 mm, and apical deviation: 1.77 ± 0.86 mm. The means for cast analyses were: angle deviation: $2.43 \pm 1.13^\circ$, coronal deviation: 0.82 ± 0.44 mm, apical deviation: 1.19 ± 0.46 mm, and depth deviation: 0.03 ± 0.65 mm. There were statistically significant differences between the deviations of CBCT scans and cast. **CONCLUSION.** The model analysis showed lower deviation value comparing the CBCT analysis. The angle and length deviation value of the universal digital guide stent were accepted clinically. [*J Adv Prosthodont 2018;10:279-85*]

KEYWORDS: Stents; Computer-Assisted Surgery; Computer-Assisted Radiotherapy Planning

INTRODUCTION

The traditional method of placing an implant is to construct a radiological guiding stents and then converting it to a surgical guiding device after Cone-beam computed tomography (CBCT) were taken.¹ However, this traditional surgical guiding stent has complicated and inaccurate lab procedure, and difficult in placing the implant fixture as planned.² The digital surgical guide was introduced to compensate this. Scanned

image of intraoral cavity and CBCT image are used to plan the placement of implant considering bone, mucosa, and tooth.³ Using digital surgical guide, drilling and placing implant at a preset position is possible, which makes less error compare to the traditional method, but only when surgical guide is maintained accurate and stable.⁴

The accuracy of implant placement with a digital surgical guide is evaluated by the deviation in the planned implant and the placed implant.⁵ A previous study showed coronal deviation of 1.09 mm, apical deviation of 1.28 mm, and axis angle deviation of 3.9° . The deviation may vary in different studies.⁶ Most previous studies were conducted with the full edentulous ridges. There are little studies about the accuracy test in the partial edentulous ridges. Furthermore, the accuracy studies were conducted with the digital guide stent only for the only one company. Also, many studies were conducted in the laboratory study, there is no clinical study. In previous studies, the accuracy of implant was assessed by overlapping CBCT before and after surgery. Analysis of implant error is not accurate due to resolution and distortion of CBCT, and error in superimposing two CBCT images. In addition, resolution is decreased due to the metal artifact when there are many metal structures.^{7,8}

Corresponding author:

Kwidug Yun

Department of Prosthodontics, School of Dentistry, Chonnam National University, 33 Youngbongro, Bukgu, Gwangju 61186, Republic of Korea
Tel. +82625305631; e-mail, ykd@jnu.ac.kr

Received October 24, 2017 / Last Revision February 1, 2018 / Accepted February 27, 2018

© 2018 The Korean Academy of Prosthodontics

This is an Open Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<http://creativecommons.org/licenses/by-nc/3.0>) which permits unrestricted non-commercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

This research was financially supported by the Ministry of Trade, Industry, and Energy (MOTIE), Korea, under the "Regional industry based organization support program" (reference number R0004032) supervised by the Korea Institute for Advancement of Technology (KIAT).

The primary purpose of this study is to assess the implant placement error by using CBCT and plaster cast after placing implant in the posterior tooth with universal digital surgical guide and kit.

MATERIALS AND METHODS

This study was approved by the Institutional Review Board, Chonnam National University Dental Hospital (IRB No. CNUDH-2016-007). To calculate the number of subjects required for this study, *in vitro* experiment⁹ was performed using a partial edentulous epoxy model (M. Tech, Seoul, Korea). The number of the subjects is 26, which is calculated using G*power 3.1 program (Heinrich Heine University, Düsseldorf, Germany). 28 implants were selected in this study considering 10% failure rate. The following criteria

were used to recruit 28 implants placement (Table 1).

Before placing an implant, patient's preliminary impression was taken and the diagnostic model was fabricated using hard plaster. 3D model scanner (Freedom HD, Degree of Freedom, Seoul, Korea) was used to scan the diagnostic model and the information of patient's intraoral soft tissue surface was saved as Surface Tessellation Language (STL) file. Patient's hard tissue information was obtained by taking CBCT (Alphard-3030, ASAHI Roentgen, Kyoto, Japan) and saved as Digital imaging and communications in medicine (DICOM) file. After superimposing the STL and DICOM files on the remaining natural teeth using In2guide (Cybermed, Seoul, Korea) software, the surgical guide was fabricated using 3D printer considering the diameter, length, and position of implant (Fig. 1). 4 weeks after placing the implant with digital surgical guide, CBCT was taken to eval-

Table 1. Patient selection criteria for inclusion and exclusion

Inclusion criteria	Exclusion criteria
① Missing premolar or molar for more than 3 months	① Acute periodontitis
② Wish to replace the missing tooth with an implant	② Heavy smoking
③ Willing to sign for informed consent	③ Physical or psychological disorders prohibiting implant treatment
④ Sufficient bone for implant placement	④ Previous therapeutic radiation of the head-neck region
⑤ Sufficient attached mucosa present	⑤ Younger than 18 years
⑥ Mouth opening \geq 40 mm	
⑦ Good general health	




Fig. 1. Pre-surgical procedure. (A) Diagnostic cast, (B) STL data scanned by 3D model scanner, (C) Pre-surgical CBCT image, (D) Planning on position and angulation of the implant, (E) Surgical guide fabricated by stereolithography.

uate the implant accuracy (Fig. 2). The term ‘planned implant’ is used to describe the preset position of implant before the surgery, and the ‘inserted implant’ is the actual placement of fixture after the surgery. The coronal deviation (in mm), apical deviation (in mm), and angle deviation (in °) of planned implant and inserted implant were measured to evaluate the accuracy (Fig. 3).

Eight weeks after the implant placement, polyvinylsiloxane (Honigum, DMG, Hamburg, Germany) was used to take impression to make hard plaster model and be scanned; this was called inserted model (Fig. 4). Lab analog inside the plaster was removed after scanning. A lab cylinder screw, a lab cylinder body, and a lab analog were connected to the metal sleeve of the surgical guide used for implant placement, and the lab analog was fixed with a hard plaster to regenerate the planned model, the model of planned position of implant before the surgery. The STL file obtained

by scanning pre-operative and postoperative models was superimposed and analyzed with a 3D analysis program (Geomagic control X, 3D Systems, Morrisville, NC, USA). Depth deviation is obtained by subtracting the z-axis coordinate of the post-operative model from the pre-operative model. If the value is positive, the implant fixture is located more apically and if the value is negative, the implant is located more on coronal (Fig. 5).

All statistical analysis is performed by SPSS Ver 23.0 (SPSS Inc., Chicago, IL, USA). Angle deviation, coronal deviation, and apical deviation measured by CBCT and cast model are statistically analyzed by Independent t-test after performing normality test using Kolmogorov-Smirnov test. All results were tested for significance at the level of $P < .05$. Pearson correlation analysis on angle, coronal and apical deviation was also used in searching for the correlativity of the two methods.




Fig. 2. Surgical procedures of flapless implant fixture installation by using surgical guide. (A) Surgical template applied over the edentulous area and adjacent teeth, (B) Removal of soft tissue with punch, (C) Drilling through the implant template with a drill, (D) Postoperative view of installed implant.




Fig. 3. Overlapping observation between planned and inserted implants for measurement of deviations. (A) Cross-sectional view of the overlapping, (B) Panoramic view of the overlapping.




Fig. 4. Inserted model. (A) The cast of inserted model, (B) STL data of inserted model scanned by 3D model scanner, (C) Component used for fabrication of planned model, (D) The cast of planned model, (E) STL data of planned model scanned by 3D model scanner.




Fig. 5. (A) Illustrations of the deviations between planned and inserted implant on the method with cast, (B) The accuracy analysis between planned and inserted implant by 3D analysis program.

RESULTS

The accuracy measured with CBCT and cast model is as follows (Table 2). The means for CBCT analyses were: angle deviation: $4.74 \pm 2.06^\circ$, coronal deviation: 1.37 ± 0.80 mm, and apical deviation: 1.77 ± 0.86 mm. The means for cast analyses were: angle deviation: $2.43 \pm 1.13^\circ$, coronal deviation: 0.82 ± 0.44 mm, apical deviation: 1.19 ± 0.46 mm, and

depth deviation: 0.03 ± 0.65 mm. The angular, coronal, and apical deviations were significantly smaller in measurement using cast model than those measured using the CBCT method ($P = .01$). Also, apical deviation was bigger than coronal deviation (Fig. 6). Angular, coronal and apical deviation in CBCT and cast showed positive correlation and there were significant difference among them ($P = .001, .024$) (Fig. 7).

Table 2. The angle, coronal, apical and depth deviation between CBCT and cast

Method		Mean ± SD	Max	Min	P value
Angle deviation (°)	CBCT	4.74 ± 2.06 ^a	8.86	0.00	0.001
	Cast	2.43 ± 1.13 ^b	5.70	1.15	
Coronal deviation (mm)	CBCT	1.37 ± 0.80 ^a	3.76	0.18	0.002
	Cast	0.82 ± 0.44 ^b	1.85	0.13	
Apical deviation (mm)	CBCT	1.77 ± 0.86 ^a	3.76	0.45	0.003
	Cast	1.19 ± 0.46 ^b	2.51	0.37	
Depth deviation (mm)	Cast	-0.03 ± 0.65	1.58	-2.12	

Different superscript letters indicate significant differences.




Fig. 6. The angle and length deviation between CBCT and cast. Results show mean ± SD angle deviation. * indicates statistical differences ($P < .05$).




Fig. 7. Scatter plot for the Pearson correlation between CBCT analysis and cast analysis. (A) The angle deviation ($P = .001$, $R = 0.635$), (B) The coronal deviation ($P = .054$, $R = 0.368$), (C) The apical deviation ($P = .024$, $R = 0.426$).

DISCUSSION

While digital surgical guide makes it possible to drill and place an implant in a preset position, the procedure has to include accurate position and accurate analysis before the surgery. Effort has been made to improve the accuracy of digital surgical guide as well as maintenance of surgical guide and errors occurring during the manufacturing process.

During implant procedure, 1 angle deviation makes 0.34 mm length deviation in the 10-mm fixture apical area. 5° angle deviation makes 1.7 mm length deviation. If the space between implant and tooth root were set to 1.5 mm during implant planning, 5° angle error will impair the tooth root. Thus the angle deviation should be no more than 3° to implant installed safely without the tooth damaged.¹⁰ If the important anatomical structure such as inferior alveolar nerve is close by, acceptable surgical guide's maximum angle deviation is less than 3° and maximum vertical error is less than -1.5 mm.¹⁰ Less loosening of implant and passive fit is possible when the angle between the hex of fixture and hexagonal freedom of abutment is less than 5° and the distance is 150 µm.^{11,12} This study shows that angular, coronal, and apical deviation are accurate enough to avoid the damage of major anatomical structure during the procedure, but less accurate to connect directly to pre-manufactured hexagonal implant prosthesis. Therefore, it is necessary to manufacture prosthesis by taking an impression after the implant placement, or to use non-hexagonal implant fixture.¹

In this study, deviation measured by CBCT is similar to that of other studies but angle deviation is somewhat higher. The reason for this is that the previous studies used surgical kit and implant fixture of the subsidiary company that makes surgical guide, while universal surgical guide kit is used in this study. Also, compared the previous studies, the more rearmost molars are included in this study. Reference marker was not used when taking CBCT, so higher error occurred during overlapping preoperative and postoperative CBCT.

The angle, coronal, and apical deviation were statistically significantly smaller than those of CBCT in this study ($P < .05$). Not only there is no error occurred while superimposing pre-operative and postoperative CBCTs, but also can additional radiation be decreased by using cast model to analyze accuracy. However, errors can occur when taking impression or making plaster cast model. There was more error in overlapping CBCTs because of the absence of marker in CBCT analysis. Error can be decreased by using only one cast model to reproduce preoperative and postoperative cast model. Further studies in comparing these two methods are required by adding marker to improve accuracy of superimposition. In cast model analysis, depth deviation can be measured by setting axis of the implant as z-axis and obtaining difference in the z-axis. Average of depth deviation is -0.03 ± 0.65 mm, which is more on the coronal side than planned position. This result is similar to that of the previous study, especially when those who have less experience with digital guided surgery tend to have less reliability

in the accuracy of the surgical guide.¹³

This study shows that the deviation obtained by the plaster cast is significantly smaller, which can be useful in evaluation of implant placement accuracy. In addition, the angular deviation may become larger toward the farthest tooth in implant placement.

CONCLUSION

The present study results showed significantly smaller deviation values using cast model analysis than those measured using the CBCT superimposition method. The angle and length deviation value of the universal digital guide stent were clinically acceptable.

ORCID

Boyoung Ma <https://orcid.org/0000-0002-8147-056X>
 Taewoek Park <https://orcid.org/0000-0003-0180-4409>
 Inkon Chun <https://orcid.org/0000-0003-0539-8313>
 Kwidug Yun <https://orcid.org/0000-0002-2965-3967>

REFERENCES

- Misch CE. Dental implant prosthetics. Elsevier Health Sciences; MO, USA, 2014.
- Nickenig HJ, Eitner S. Reliability of implant placement after virtual planning of implant positions using cone beam CT data and surgical (guide) templates. *J Craniomaxillofac Surg* 2007;35:207-11.
- Ganz SD. Three-dimensional imaging and guided surgery for dental implants. *Dent Clin North Am* 2015;59:265-90.
- Shim JS, Kim NH, Kim JE. A procedure for the computer-guided implant planning: A narrative review. *J Korean Dent Assoc* 2016;54:108-22.
- Van Assche N, Vercruyssen M, Coucke W, Teughels W, Jacobs R, Quirynen M. Accuracy of computer-aided implant placement. *Clin Oral Implants Res* 2012;23:112-23.
- Sicilia A, Botticelli D; Working Group 3. Computer-guided implant therapy and soft- and hard-tissue aspects. The third EAO consensus conference 2012. *Clin Oral Implants Res* 2012;23:157-61.
- Martorelli M. A new approach in CT artifact removal: three cases study in maxillofacial surgery. *Int J Interact Des Manuf* 2013;7:115-24.
- Komiyama A, Pettersson A, Hultin M, Näsström K, Klinge B. Virtually planned and template-guided implant surgery: an experimental model matching approach. *Clin Oral Implants Res* 2011;22:308-13.
- Yoon JH. The accuracy estimate of surgical stents fabricated by digital methods in installing implants on dental models. Chonnam National University. Master Degree Thesis. 2017.
- Choi B, Jeong S. Digital flapless implantology. Seoul; Ji-Sung Publishing Co.; 2015. p. 32-51.
- Al Quran FA, Rashdan BA, Zomar AA, Weiner S. Passive fit and accuracy of three dental implant impression techniques. *Quintessence Int* 2012;43:119-25.

12. Binon PP. The effect of implant/abutment hexagonal misfit on screw joint stability. *Int J Prosthodont* 1996;9:149-60.
13. Pozzi A, Polizzi G, Moy PK. Guided surgery with tooth-supported templates for single missing teeth: A critical review. *Eur J Oral Implantol* 2016;9:S135-53.
14. Verhamme LM, Meijer GJ, Boumans T, de Haan AF, Bergé SJ, Maal TJ. A clinically relevant accuracy study of computer-planned implant placement in the edentulous maxilla using mucosa-supported surgical templates. *Clin Implant Dent Relat Res* 2015;17:343-52.
15. Ersoy AE, Turkyilmaz I, Ozan O, McGlumphy EA. Reliability of implant placement with stereolithographic surgical guides generated from computed tomography: clinical data from 94 implants. *J Periodontol* 2008;79:1339-45.
16. Behneke A, Burwinkel M, Behneke N. Factors influencing transfer accuracy of cone beam CT-derived template-based implant placement. *Clin Oral Implants Res* 2012;23:416-23.
17. Kang BG, Kim HJ, Chung CH. Accuracy of the CT guided implant template by using an intraoral scanner according to the edentulous distance. *J Korean Acad Prosthodont* 2017;55: 1-8.
18. Pettersson A, Kero T, Gillot L, Cannas B, Fäldt J, Söderberg R, Näsström K. Accuracy of CAD/CAM-guided surgical template implant surgery on human cadavers: Part I. *J Prosthet Dent* 2010;103:334-42.
19. Cassetta M, Di Mambro A, Giansanti M, Stefanelli LV, Cavallini C. The intrinsic error of a stereolithographic surgical template in implant guided surgery. *Int J Oral Maxillofac Surg* 2013;42:264-75.
20. Ozan O, Turkyilmaz I, Ersoy AE, McGlumphy EA, Rosenstiel SF. Clinical accuracy of 3 different types of computed tomography-derived stereolithographic surgical guides in implant placement. *J Oral Maxillofac Surg* 2009;67:394-401.
21. Valente F, Schiroli G, Sbrenna A. Accuracy of computer-aided oral implant surgery: a clinical and radiographic study. *Int J Oral Maxillofac Implants* 2009;24:234-42.
22. Vasak C, Watzak G, Gahleitner A, Strbac G, Schemper M, Zechner W. Computed tomography-based evaluation of template (NobelGuide™)-guided implant positions: a prospective radiological study. *Clin Oral Implants Res* 2011;22:1157-63.
23. Park C, Raigrodski AJ, Rosen J, Spiekerman C, London RM. Accuracy of implant placement using precision surgical guides with varying occlusogingival heights: an in vitro study. *J Prosthet Dent* 2009;101:372-81.