

S1 Table. CT techniques

	CT parameter
Beam collimation (mm)	64×0.6
Beam pitch	1
Gantry rotation time (s)	0.5
kV/mAs	120/200
Reconstruction thickness	
Axial AP/PVP/DP (mm)	5/3/3
Coronal AP/PVP/DP (mm)	5/5/5
Delay time for AP (s)	15

AP, arterial phase; CT, computed tomography; DP, delayed phase; PVP, portal venous phase.