

Table S2. Interfering substances used in the specificity test of BRCAaccuTest™

Interfering substance	Normal range	Addition amount	Manufacturer
Bilirubin	0.2-1.2 mg/dL in blood	40 mg/dL	Sigma Aldrich
Hemoglobin	12-17 g/dL in blood	2 g/dL	Sigma Aldrich
Cholesterol	<200 mg/dL in blood	503 mg/dL	Sigma Aldrich
Wash buffer	0% in DNA sample	15% in total volume	Qiagen