



Supplementary Fig. 2. Correlation between the reliability indicator and accuracy of the measurements at varying stiffness values. Scatter-plot analysis of all measurements, plotted as percentage deviations from the reference stiffness values compared with the reliability indicator. All trendlines had R^2 values of <0.5 , indicating a poor correlation between the two variables. For 19.2 kPa, an R^2 value for EUS-SWE was not reported, as all values had a VsN of 100%. VsN, reliability indicator; EUS-SWE, endoscopic ultrasound shear wave elastography.