

**Supplemental Data Table S2.** Precision evaluation of the ID-LC–MS/MS method

Sample	Mean (pmol/L)	Within-run CV (%)	Between-run CV (%)	Total CV (%)
Concentration 1	199.3	2.1	0.6	2.2
Concentration 2	634.4	1.0	0.9	1.3
Concentration 3	1,573.1	1.1	0.8	1.3
Concentration 4	2,836.8	1.0	1.2	1.6

Abbreviation: ID-LC–MS/MS, isotope dilution liquid chromatography-tandem mass spectrometry.