

**Supplemental Data Table S2.** Analytical performance specifications based on the BV of the measurand

Test	Analytical specifications for total error		
	Minimum	Desirable	Optimum
HbA1c	3.3	2.2	1.1
Cr	11.1	7.4	3.7
TC	13.0	8.7	4.3
TG	40.6	27.0	13.5
AFP	52.2	34.8	17.4
PSA	24.4	16.2	8.1
cTnI	35.3	23.5	11.8

Abbreviations: BV, biological variation; Cr, creatinine; TC, total cholesterol; TG, triglyceride; AFP, alpha-fetoprotein; PSA, prostate-specific antigen; cTnI, cardiac troponin I.